



**SPORTON LAB.**

Certificate No: **EC8N2719**

# CERTIFICATE



**EQUIPMENT : Digital Multimeter**

**MODEL NO. : BM115, BM116, BM117, BM118**

**APPLICANT : BRYMEN TECHNOLOGY CORPORATION.**

7F, No. 758, Chung Jeng Road,

Chung Ho City 235, Taipei, Taiwan, R.O.C.



**I HEREBY**

**CERTIFY THAT:**

THE MEASUREMENTS SHOWN IN THIS TEST REPORT WERE MADE IN ACCORDANCE WITH THE PROCEDURES GIVEN IN **EUROPEAN COUNCIL DIRECTIVE 2004/108/EC**. THE EQUIPMENT WAS **PASSED** THE TEST PERFORMED ACCORDING TO  
**European Standard EN 61326-1:2006,**  
**( EN 55011:1998/A1:1999/A2:2002 Class B,**  
**EN 61000-3-2:2006, EN 61000-3-3:1995/A1:2001/A2:2005,**  
**IEC 61000-4-2:1995/A1:1998/A2:2000, IEC 61000-4-3:2002,**  
**IEC 61000-4-4:2004, IEC 61000-4-5:1995/A1:2000, IEC 61000-4-6:2003,**  
**IEC 61000-4-8:1993/A1:2000, IEC 61000-4-11:2004 ).**  
THE TEST WAS CARRIED OUT ON **Dec. 08, 2008** AT **SPORTON INTERNATIONAL INC. LAB.**

*Castries Huang Dec 08, 2008*  
Castries Huang  
Supervisor